

Quick Reference Guide

Reference Material



manual.carelyteplus.com

Set Up, Installation, Preparation, Calibration, Components, Language Selection, Shutdown





manual.carelyteplus.com

Maintenance & Troubleshooting, Mechanical Assemblies, Adjustments, Software Operation





ideo.carelyteplus.com

Online Training, Maintenance & Troubleshooting, Setup and Quick Reference Videos





Quick Reference Guide

Measurement

















Always wear gloves when handling blood or parts potentially contaminated with blood.





Keep the main door closed during measurement.

Maintenance	
Daily	 Perform cleaning cycle at end of the day Perform conditioning cycle at end of the day Check printer paper supply
Weekly	Clean sample probe and fill portClean exterior analyzer surfaces
Monthly	 Clean the reference electrode housing (see chapter 5 of Operator Manual)
Semi-annually	Replace peristaltic pump tubing
Annually	Replace main tubing harness

Troubleshooting

- Open front cover and check if all electrodes and the right side sample sensor have O-rings.
- Make sure maintenance has been performed in accordance with the maintenance schedule.
- Check the FLUID PACK level by going to menu [PACK STATUS]
- Check for clots and crystals that may be forming in the tubing or in the measuring chamber.
- Pull pump windings from pump roller and check for kinked tubing and tears in the tubing.
- Start a calibration pressing [CALIBRATION]